



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Northwell Health Laboratories**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Roche Cobas c513	Calibrator Lot: 74039400
Reagent Lot: 804154, 863606, 867833, 849213, 860214	Calibrator Assigned Values: 0.224, 0.562, 1.1, 1.52, 2.096, 2.98 mmol/L HbA1c; 2.4, 12.91 mmol/L Hb

Date of Certification: September 1, 2025

Certification Expires: September 1, 2026

A handwritten signature in black ink, appearing to read 'David Sacks'.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Curt Rohlfing'.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'Rhonda Haward'.

SRL director/ supervisor