



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Northwell Health Laboratories**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Roche Cobas c513	Calibrator Lot: 8008110
Reagent Lot: 926424, 938276, 928774, 943150, 942411	Calibrator Assigned Values: 0.238, 0.536, 1.011, 1.38, 1.892, 2.669 g/dL HbA1c; 2.38, 12.82 g/dL

Date of Certification: September 1, 2026

Certification Expires: September 1, 2027

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Curt Rohlfing".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to be a stylized monogram.

SRL director/ supervisor